

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1017

Name: RITTER 1  
Sterilizer ID: V2178198  
07/15/2021 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
2:00 184.9 F 3.7  
4:00 193.8 F 7.7  
6:00 217.4 F 12.9  
8:00 227.2 F 6.8  
10:00 238.9 F 11.9  
12:00 237.0 F 9.1  
14:00 239.4 F 9.7  
16:00 248.0 F 14.0  
18:00 260.0 F 20.5  
20:00 270.7 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 272.5 F 28.7  
2:00 271.8 F 28.1  
3:00 271.4 F 28.3  
4:00 272.4 F 28.7

Min 271.1 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
5:00 228.3 F 0.0  
10:00 226.1 F 0.0  
15:00 227.6 F 0.0  
20:00 232.6 F 0.0  
25:00 231.0 F 0.0  
30:00 211.2 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 20:51  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

07/15/2021 11:01 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02