
Midmark M11 - v1.0.0
Total Cycles: 4894

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2024 04:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.1 F 0.0
2:00 135.6 F 0.1
4:00 147.5 F 0.3
6:00 175.8 F 0.6
8:00 196.5 F 2.3
10:00 211.6 F 5.7
12:00 221.1 F 9.0
14:00 234.4 F 12.7
16:00 233.0 F 8.5
18:00 243.3 F 13.2
20:00 241.1 F 11.0
22:00 237.8 F 9.3
24:00 250.9 F 15.4
26:00 261.1 F 21.2
28:00 270.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 271.6 F 28.3
2:00 271.9 F 28.5
3:00 272.3 F 28.8
4:00 272.4 F 28.9

Min 271.2 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 228.1 F 0.0
10:00 241.3 F 0.0
15:00 238.5 F 0.0
20:00 226.1 F 0.0
25:00 226.6 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:07
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:07:55
CYCLE COMPLETE

09/11/2024 05:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____