
Midmark M11 - v1.0.5
Total Cycles: 607

Name: RITTER2
Sterilizer ID: V2313228
12/07/2021 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	71.9 F	0.0
2:00	82.9 F	0.0
4:00	114.7 F	1.1
6:00	147.8 F	3.9
8:00	178.7 F	7.9
10:00	200.8 F	12.6
12:00	212.6 F	4.9
14:00	225.4 F	9.3
16:00	232.5 F	9.8
18:00	238.7 F	10.9
20:00	235.3 F	8.6
22:00	250.9 F	16.0
24:00	263.9 F	23.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.8 F	28.8
1:00	272.7 F	29.6
2:00	272.4 F	28.9
3:00	271.9 F	29.0
4:00	272.9 F	29.5

Min 271.8 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	192.3 F	0.0
5:00	174.5 F	0.0
10:00	136.9 F	0.0
15:00	125.4 F	0.0

Dry Cycle Aborted

FILLING: 1:18
HEATING: 25:46
STERILIZING: 4:00
VENTING: 2:24
DRYING: 17:17
TOTAL CYCLE: 00:50:44

12/07/2021 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____