
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001805
10:04 26/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:04
STERILIZING 3:03
136.0 C 220kPa 3:03
136.1 C 223kPa 3:33
136.4 C 224kPa 4:03
136.3 C 222kPa 4:33
136.4 C 224kPa 5:03
136.4 C 226kPa 5:33
136.3 C 224kPa 6:03
136.3 C 224kPa 6:33
136.3 C 224kPa 6:33
VENTING 6:33
AIR DRYING 7:24
CYCLE COMPLETE 37:24

Digital Signature #
269ADADC6EF3C8F7
