
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000711
14:28 28/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:04
136.0 C 220kPa 3:04
136.3 C 217kPa 3:35
136.3 C 224kPa 4:05
136.2 C 223kPa 4:35
136.1 C 221kPa 5:05
136.1 C 221kPa 5:35
136.2 C 220kPa 6:05
136.1 C 222kPa 6:35
136.1 C 222kPa 6:35
VENTING 6:35
AIR DRYING 7:28
CYCLE COMPLETE 7:33

Digital Signature #
5A98F443CA39A168
