
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000167
8:14 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:14
PRESSURIZING 4:14
STERILIZING 5:39
136.5 C 225kPa 5:39
136.4 C 227kPa 6:09
136.4 C 226kPa 6:39
136.3 C 223kPa 7:09
136.4 C 226kPa 7:39
136.3 C 225kPa 8:09
136.4 C 226kPa 8:39
136.4 C 226kPa 9:09
136.4 C 226kPa 9:09
VENTING 9:09
AIR DRYING 10:13
CYCLE COMPLETE 25:13

Digital Signature #
4A557821548A2AA7
