
Midmark M11 - v1.0.0
Total Cycles: 3007

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2023 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 198.8 F 3.2
4:00 201.4 F 7.8
6:00 214.7 F 13.3
8:00 229.5 F 8.7
10:00 240.0 F 10.7
12:00 241.6 F 11.9
14:00 240.0 F 10.4
16:00 253.8 F 17.3
18:00 265.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.0
2:00 271.7 F 28.3
3:00 272.4 F 29.2
4:00 272.3 F 28.8

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 225.4 F 0.0
10:00 231.9 F 0.0
15:00 230.1 F 0.0
20:00 227.5 F 0.0
25:00 227.0 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

03/15/2023 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____