
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002421
11:11 14/01/2025
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 5:19
STERILIZING 6:26
136.4 C 225kPa 6:26
136.2 C 223kPa 6:56
136.2 C 223kPa 7:26
136.3 C 224kPa 7:56
136.3 C 224kPa 8:26
136.3 C 223kPa 8:56
136.1 C 221kPa 9:26
136.4 C 225kPa 9:56
136.1 C 222kPa 10:26
136.4 C 225kPa 10:56
136.4 C 226kPa 11:26
136.3 C 224kPa 11:56
136.3 C 225kPa 12:26
136.4 C 226kPa 12:56
136.2 C 223kPa 13:26
136.4 C 225kPa 13:56
136.4 C 224kPa 14:26
136.4 C 223kPa 14:56
136.1 C 223kPa 15:26
136.2 C 225kPa 15:56
136.2 C 224kPa 16:26
136.2 C 224kPa 16:26
VENTING 16:26
AIR DRYING 17:19
CYCLE COMPLETE 17:38

Digital Signature #
823E5EC8D90B1C20
