

-----  
Midmark M11 - v1.0.0  
Total Cycles: 727

Name: RITTER 1  
Sterilizer ID: V2178198  
03/17/2021 02:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.0 F 0.0  
2:00 153.4 F 1.9  
4:00 163.7 F 5.8  
6:00 201.7 F 10.5  
8:00 222.5 F 4.5  
10:00 226.8 F 10.3  
12:00 232.9 F 7.2  
14:00 245.2 F 13.4  
16:00 243.3 F 11.9  
18:00 256.7 F 18.8  
20:00 268.3 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.1 F 28.5  
2:00 272.7 F 28.8  
3:00 272.6 F 29.2  
4:00 271.4 F 28.5  
  
Min 271.2 F 28.4  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.3 F 0.0  
5:00 241.5 F 0.0  
10:00 213.6 F 0.0  
15:00 213.3 F 0.0  
20:00 217.6 F 0.0  
25:00 181.5 F 0.0  
30:00 185.3 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 21:07  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 00:59:15  
CYCLE COMPLETE

03/17/2021 03:53 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01