

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007150  
9:27 05/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:28  
STERILIZING 3:22  
136.0 C 220kPa 3:22  
136.3 C 221kPa 3:52  
136.2 C 223kPa 4:22  
136.4 C 220kPa 4:52  
136.3 C 223kPa 5:22  
136.1 C 219kPa 5:52  
136.2 C 224kPa 6:22  
136.4 C 220kPa 6:52  
136.4 C 220kPa 6:52  
VENTING 6:52  
AIR DRYING 7:42  
CYCLE COMPLETE 17:06

-----  
Digital Signature #  
C6238123484AA618  
-----