
Midmark M11 - v1.0.0
Total Cycles: 4951

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2024 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 73.7 F 0.0
4:00 85.5 F 0.1
6:00 111.6 F 0.1
8:00 150.4 F 0.6
10:00 176.7 F 2.7
12:00 203.0 F 5.7
14:00 209.8 F 9.0
16:00 222.2 F 12.5
18:00 222.8 F 5.3
20:00 232.9 F 9.3
22:00 242.1 F 13.5
24:00 239.2 F 10.2
26:00 236.1 F 8.2
28:00 248.0 F 14.0
30:00 259.1 F 20.0
32:00 269.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.7 F 28.8
4:00 272.0 F 28.2

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 221.8 F 0.0
10:00 207.1 F 0.0
15:00 205.8 F 0.0
20:00 187.7 F 0.0
25:00 190.7 F 0.0
30:00 191.9 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 33:08
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:11:12
CYCLE COMPLETE

10/02/2024 09:37 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____