
Midmark M11 - v1.0.0
Total Cycles: 4984

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2024 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.3 F 0.0
2:00 173.3 F 0.5
4:00 176.3 F 1.6
6:00 194.5 F 5.7
8:00 210.7 F 9.8
10:00 226.5 F 9.0
12:00 227.3 F 8.5
14:00 238.2 F 13.3
16:00 238.6 F 10.6
18:00 235.8 F 8.4
20:00 249.7 F 15.1
22:00 261.0 F 21.4
24:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 29.0
2:00 272.0 F 28.6
3:00 271.3 F 28.1
4:00 271.8 F 28.8

Min 271.2 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 206.1 F 0.0
10:00 205.9 F 0.0
15:00 209.3 F 0.0
20:00 215.2 F 0.0
25:00 220.2 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:45
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:29
CYCLE COMPLETE

10/10/2024 10:37 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____