
Midmark M11 - v1.0.0
Total Cycles: 1263

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2021 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 91.2 F 0.0
4:00 90.1 F 0.0
6:00 118.1 F 0.1
8:00 154.4 F 0.1
10:00 184.9 F 0.2
12:00 209.8 F 1.7
14:00 226.6 F 5.0
16:00 233.8 F 7.8
18:00 241.5 F 11.3
20:00 233.0 F 7.3
22:00 246.4 F 13.1
24:00 242.7 F 11.3
26:00 239.9 F 10.1
28:00 254.0 F 17.0
30:00 265.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.4 F 28.3
2:00 272.3 F 28.5
3:00 272.0 F 28.0
4:00 272.6 F 28.7

Min 271.5 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 196.9 F 0.0
10:00 195.6 F 0.0
15:00 189.7 F 0.0
20:00 225.8 F 0.0
25:00 230.5 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:41
CYCLE COMPLETE

10/14/2021 08:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____