
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002189
8:49 01/11/2024
AIR DRYING ONLY
60 min. MAXIMUM
CYCLE START 0:00
AIR DRYING 0:02
CYCLE COMPLETE 7:40

Digital Signature #
E0D966D9FEC4B590
