

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003593  
14:42 03/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:32  
STERILIZING 3:37  
136.5 C 220kPa 3:37  
136.2 C 223kPa 4:07  
136.2 C 223kPa 4:37  
136.2 C 224kPa 5:07  
136.4 C 222kPa 5:37  
136.4 C 224kPa 6:07  
136.3 C 224kPa 6:37  
136.2 C 223kPa 7:07  
136.2 C 223kPa 7:07  
VENTING 7:07  
AIR DRYING 7:58  
CYCLE COMPLETE 8:14  
-----

Digital Signature #  
B9405CF8801AD5BF  
-----