

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1210

Name: RITTER 1  
Sterilizer ID: V2178198  
09/28/2021 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.3 F 0.0  
2:00 109.9 F 0.0  
4:00 137.8 F 0.1  
6:00 154.2 F 0.2  
8:00 204.8 F 0.3  
10:00 209.3 F 3.6  
12:00 218.9 F 6.6  
14:00 232.0 F 10.1  
16:00 234.2 F 9.2  
18:00 237.1 F 10.3  
20:00 235.9 F 8.7  
22:00 246.1 F 13.0  
24:00 243.0 F 11.4  
26:00 255.1 F 17.6  
28:00 265.7 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.9  
1:00 272.7 F 28.6  
2:00 271.7 F 27.9  
3:00 272.0 F 28.5  
4:00 272.4 F 28.4

Min 271.2 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 231.1 F 0.0  
10:00 221.0 F 0.0  
15:00 218.1 F 0.0  
20:00 230.4 F 0.0  
25:00 220.1 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 29:50  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:08:01  
CYCLE COMPLETE

09/28/2021 09:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01