
Midmark M11 - v1.0.0
Total Cycles: 2105

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2022 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.2 F 0.0
2:00 141.7 F 0.3
4:00 149.6 F 2.6
6:00 181.0 F 5.6
8:00 206.2 F 10.3
10:00 220.0 F 4.7
12:00 227.9 F 8.7
14:00 238.5 F 13.4
16:00 238.3 F 10.2
18:00 236.4 F 8.3
20:00 248.1 F 14.4
22:00 259.6 F 20.7
24:00 270.2 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 271.5 F 28.6
2:00 272.6 F 29.2
3:00 272.1 F 28.6
4:00 271.3 F 28.3

Min 271.2 F 28.2
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 208.6 F 0.0
10:00 206.0 F 0.0
15:00 205.9 F 0.0
20:00 209.9 F 0.0
25:00 208.8 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:04
CYCLE COMPLETE

07/08/2022 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____