
Midmark M11 - v1.0.0
Total Cycles: 3179

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2023 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 182.5 F 0.5
4:00 184.9 F 3.6
6:00 209.6 F 8.5
8:00 228.6 F 10.1
10:00 232.2 F 9.3
12:00 239.1 F 9.5
14:00 244.1 F 13.0
16:00 242.6 F 12.0
18:00 256.0 F 18.9
20:00 267.7 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.3 F 29.2
2:00 271.3 F 28.7
3:00 272.4 F 29.2
4:00 271.3 F 28.5

Min 271.2 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 217.0 F 0.0
10:00 240.6 F 0.0
15:00 234.8 F 0.0
20:00 228.3 F 0.0
25:00 228.3 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

05/01/2023 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____