

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1344

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2022 07:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.6 F 0.0  
2:00 175.8 F 0.0  
4:00 167.0 F 1.1  
6:00 201.1 F 5.7  
8:00 222.9 F 10.7  
10:00 226.1 F 5.6  
12:00 239.3 F 11.1  
14:00 234.8 F 8.6  
16:00 241.5 F 10.7  
18:00 246.8 F 13.8  
20:00 258.8 F 20.5  
22:00 269.8 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.4 F 29.3  
2:00 271.5 F 29.0  
3:00 272.3 F 29.2  
4:00 271.4 F 28.8

Min 271.4 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 233.8 F 0.0  
10:00 213.0 F 0.0  
15:00 234.9 F 0.0  
20:00 222.1 F 0.0  
25:00 223.0 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 22:59  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:34  
CYCLE COMPLETE

07/26/2022 08:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_