
Midmark M11 - v1.0.0
Total Cycles: 1476

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2021 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.3 F 0.0
2:00 132.3 F 0.0
4:00 152.8 F 0.0
6:00 183.7 F 0.3
8:00 200.2 F 1.2
10:00 214.8 F 4.5
12:00 227.5 F 7.7
14:00 239.1 F 11.0
16:00 235.0 F 8.0
18:00 244.2 F 12.4
20:00 240.8 F 10.5
22:00 237.3 F 8.9
24:00 251.5 F 15.7
26:00 262.9 F 22.2
28:00 272.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 272.8 F 28.7
2:00 272.0 F 28.1
3:00 271.6 F 27.9
4:00 272.8 F 28.7

Min 271.4 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.1 F 0.0
5:00 198.6 F 0.0
10:00 188.4 F 0.0
15:00 187.8 F 0.0
20:00 185.1 F 0.0
25:00 208.0 F 0.0
30:00 210.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:19
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:06:21
CYCLE COMPLETE

12/20/2021 02:39 PM
ACCEPTED: X
REJECTED:
Operator: U02