
Midmark M11 - v1.0.0
Total Cycles: 2381

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2022 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 89.5 F 0.0
4:00 91.8 F 0.0
6:00 129.8 F 0.1
8:00 164.0 F 0.7
10:00 182.6 F 4.6
12:00 200.9 F 8.5
14:00 213.8 F 12.4
16:00 223.5 F 5.9
18:00 234.8 F 10.2
20:00 237.5 F 8.7
22:00 244.3 F 12.3
24:00 241.5 F 10.6
26:00 255.0 F 17.4
28:00 266.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.7
1:00 271.8 F 28.0
2:00 272.3 F 28.4
3:00 272.7 F 28.5
4:00 272.7 F 28.4

Min 271.5 F 27.6
Max 272.7 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 234.2 F 0.0
5:00 227.7 F 0.0
10:00 233.0 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 29:22
STERILIZING: 4:00
VENTING: 2:47
DRYING: 12:34
TOTAL CYCLE: 00:49:35

09/29/2022 11:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____