

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2528

Name: RITTER2  
Sterilizer ID: V2313228  
06/19/2023 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.3 F 0.0  
2:00 119.7 F 0.0  
4:00 121.7 F 0.6  
6:00 152.8 F 4.0  
8:00 181.1 F 7.7  
10:00 200.9 F 12.0  
12:00 214.8 F 5.6  
14:00 228.7 F 10.6  
16:00 231.5 F 7.7  
18:00 243.1 F 13.2  
20:00 241.5 F 11.7  
22:00 253.8 F 18.0  
24:00 264.7 F 24.6  
26:00 271.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.9  
1:00 272.7 F 30.5  
2:00 271.9 F 30.0  
3:00 272.7 F 30.2  
4:00 272.3 F 30.5

Min 271.9 F 29.8  
Max 272.9 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 209.4 F 0.0  
10:00 218.9 F 0.0  
15:00 224.3 F 0.0  
20:00 227.6 F 0.0  
25:00 227.9 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 26:49  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:04:53  
CYCLE COMPLETE

06/19/2023 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_