

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000707  
8:34 20/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:54  
PRESSURIZING 3:55  
STERILIZING 5:03  
136.4 C 233kPa 5:03  
136.3 C 231kPa 5:33  
136.2 C 231kPa 6:03  
136.3 C 232kPa 6:33  
136.4 C 233kPa 7:03  
136.3 C 232kPa 7:33  
136.2 C 229kPa 8:03  
136.4 C 232kPa 8:33  
136.4 C 232kPa 8:33  
VENTING 8:33  
AIR DRYING 9:28  
CYCLE COMPLETE 9:37

-----  
Digital Signature #  
34DBA4E065B570E2  
-----