
Midmark M11 - v1.0.0
Total Cycles: 3761

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.7 F 0.0
2:00 170.9 F 0.3
4:00 206.4 F 3.6
6:00 217.5 F 7.7
8:00 229.0 F 12.3
10:00 230.7 F 7.8
12:00 243.8 F 12.8
14:00 240.7 F 10.6
16:00 237.1 F 9.0
18:00 251.3 F 15.7
20:00 262.7 F 22.3
22:00 272.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.1 F 28.7
2:00 271.2 F 28.1
3:00 272.3 F 28.8
4:00 271.6 F 28.2

Min 271.2 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.0 F 0.0
5:00 229.7 F 0.0
10:00 239.7 F 0.0
15:00 212.2 F 0.0
20:00 188.2 F 0.0
25:00 183.0 F 0.0
30:00 179.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

09/12/2023 03:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____