
Midmark M11 - v1.0.5
Total Cycles: 3139

Name: RITTER2
Sterilizer ID: V2313228
11/17/2023 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 148.4 F 1.8
4:00 162.6 F 5.2
6:00 187.3 F 9.5
8:00 210.1 F 7.8
10:00 221.3 F 9.0
12:00 234.3 F 10.4
14:00 238.3 F 11.6
16:00 237.7 F 10.3
18:00 251.2 F 16.9
20:00 262.9 F 23.7
22:00 272.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.1 F 30.2
2:00 272.5 F 29.8
3:00 271.4 F 29.6
4:00 272.7 F 30.3

Min 271.3 F 29.4
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 226.0 F 0.0
10:00 218.0 F 0.0
15:00 206.3 F 0.0
20:00 204.2 F 0.0
25:00 206.7 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

11/17/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____