

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004345  
12:06 13/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:38  
STERILIZING 3:46  
136.0 C 220kPa 3:46  
136.1 C 222kPa 4:16  
136.4 C 225kPa 4:46  
136.4 C 223kPa 5:16  
136.4 C 225kPa 5:46  
136.3 C 225kPa 6:16  
136.2 C 223kPa 6:46  
136.2 C 222kPa 7:16  
136.2 C 222kPa 7:16  
VENTING 7:16  
AIR DRYING 8:09  
CYCLE COMPLETE 8:48

-----  
Digital Signature #  
0A862EE4BFD601E3  
-----