
Midmark M11 - v1.0.5
Total Cycles: 1146

Name: RITTER2
Sterilizer ID: V2313228
05/31/2022 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 187.2 F 3.9
4:00 195.3 F 9.2
6:00 221.5 F 7.4
8:00 233.0 F 10.9
10:00 233.3 F 8.5
12:00 240.0 F 10.5
14:00 246.9 F 14.1
16:00 259.4 F 20.9
18:00 270.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.6
1:00 272.5 F 29.6
2:00 272.4 F 30.0
3:00 272.0 F 30.0
4:00 272.3 F 29.6

Min 271.8 F 29.4
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 214.5 F 0.0
10:00 229.9 F 0.0
15:00 222.9 F 0.0
20:00 220.1 F 0.0
25:00 215.6 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

05/31/2022 01:13 PM
ACCEPTED: X
REJECTED:
Operator: U02