

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2749

Name: RITTER 1  
Sterilizer ID: V2178198  
01/09/2023 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.1 F 0.0  
2:00 76.7 F 0.0  
4:00 89.8 F 0.0  
6:00 114.0 F 0.0  
8:00 145.3 F 0.0  
10:00 172.5 F 2.4  
12:00 195.9 F 5.4  
14:00 203.2 F 9.1  
16:00 216.0 F 12.7  
18:00 221.5 F 6.3  
20:00 232.4 F 10.5  
22:00 233.9 F 8.2  
24:00 242.5 F 12.2  
26:00 239.4 F 10.2  
28:00 251.9 F 16.3  
30:00 262.7 F 22.5  
32:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.9  
2:00 271.4 F 28.2  
3:00 272.5 F 29.0  
4:00 271.8 F 28.3  
  
Min 271.4 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
5:00 201.0 F 0.0  
10:00 215.2 F 0.0  
15:00 219.3 F 0.0  
20:00 216.8 F 0.0  
25:00 219.4 F 0.0  
30:00 219.2 F 0.0  
  
Drying Complete

FILLING: 0:47  
HEATING: 32:24  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:10:21  
CYCLE COMPLETE

01/09/2023 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_