

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001459  
14:15 09/29/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:29  
PRESSURIZING 6:13  
STERILIZING 7:45  
136.0 C 220kPa 7:45  
136.4 C 220kPa 8:15  
136.3 C 222kPa 8:45  
135.8 C 220kPa 9:15  
136.5 C 223kPa 9:45  
136.4 C 220kPa 10:15  
136.3 C 221kPa 10:45  
136.4 C 221kPa 11:15  
VENTING 11:15  
AIR DRYING 12:21  
CYCLE COMPLETE 42:21  
-----

Digital Signature #  
E0CF0E7858C22516  
-----