
Midmark M11 - v1.0.5
Total Cycles: 1283

Name: RITTER2
Sterilizer ID: V2313228
07/11/2022 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 181.8 F 4.0
4:00 194.1 F 8.4
6:00 217.1 F 13.4
8:00 222.0 F 9.0
10:00 233.7 F 10.3
12:00 240.3 F 12.2
14:00 239.5 F 11.0
16:00 252.9 F 17.6
18:00 264.2 F 24.4
20:00 272.0 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.8 F 30.3
2:00 271.7 F 29.6
3:00 272.7 F 30.4
4:00 272.2 F 29.8

Min 271.6 F 29.5
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 212.5 F 0.0
10:00 218.6 F 0.0
15:00 221.5 F 0.0
20:00 220.7 F 0.0
25:00 219.4 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

07/11/2022 03:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____