
Midmark M11 - v1.0.5
Total Cycles: 4077

Name: RITTER2
Sterilizer ID: V2313228
11/01/2024 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.0 F 0.0
2:00 168.8 F 3.3
4:00 190.5 F 8.4
6:00 214.5 F 8.4
8:00 228.4 F 9.2
10:00 234.8 F 9.4
12:00 244.9 F 13.3
14:00 244.4 F 12.7
16:00 258.1 F 20.1
18:00 270.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.8
1:00 272.8 F 29.7
2:00 272.2 F 28.7
3:00 272.9 F 29.7
4:00 272.2 F 29.0

Min 271.7 F 28.6
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 229.6 F 0.0
10:00 227.1 F 0.0
15:00 221.5 F 0.0
20:00 219.5 F 0.0
25:00 219.9 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 18:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

11/01/2024 04:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____