

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2513

Name: RITTER2  
Sterilizer ID: V2313228  
06/14/2023 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.3 F 0.0  
2:00 166.6 F 2.8  
4:00 179.0 F 7.0  
6:00 200.9 F 12.1  
8:00 217.9 F 6.8  
10:00 232.5 F 12.2  
12:00 234.7 F 9.8  
14:00 235.9 F 9.0  
16:00 248.5 F 15.4  
18:00 260.6 F 22.2  
20:00 271.5 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.8 F 30.3  
2:00 272.1 F 29.8  
3:00 271.2 F 29.5  
4:00 272.2 F 30.3

Min 271.1 F 29.4  
Max 273.0 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.9 F 0.0  
5:00 227.5 F 0.0  
10:00 230.4 F 0.0  
15:00 240.4 F 0.0  
20:00 223.6 F 0.0  
25:00 220.8 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:59:29  
CYCLE COMPLETE

06/14/2023 10:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_