

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2354

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2023 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
2:00 196.4 F 3.4  
4:00 205.2 F 8.9  
6:00 223.1 F 8.0  
8:00 233.0 F 11.3  
10:00 234.3 F 9.3  
12:00 237.5 F 9.2  
14:00 249.1 F 15.2  
16:00 261.6 F 22.4  
18:00 272.8 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.4 F 29.4  
2:00 272.3 F 29.7  
3:00 272.0 F 29.0  
4:00 272.7 F 29.8  
  
Min 271.6 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.8 F 0.0  
5:00 224.8 F 0.0  
10:00 235.7 F 0.0  
15:00 239.9 F 0.0  
20:00 228.4 F 0.0  
25:00 221.9 F 0.0  
30:00 217.2 F 0.0  
  
Drying Complete

FILLING: 2:33  
HEATING: 18:20  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

05/11/2023 11:10 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_