
Midmark M11 - v1.0.0
Total Cycles: 3660

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2023 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.3 F 0.0
4:00 83.3 F 0.0
6:00 106.1 F 0.2
8:00 138.7 F 2.6
10:00 167.1 F 5.4
12:00 199.1 F 8.3
14:00 201.5 F 12.3
16:00 215.8 F 4.7
18:00 226.6 F 8.4
20:00 233.9 F 12.4
22:00 233.4 F 8.5
24:00 245.2 F 13.2
26:00 242.3 F 11.5
28:00 254.2 F 17.6
30:00 264.6 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.9 F 28.5
2:00 272.3 F 28.9
3:00 273.0 F 29.2
4:00 271.9 F 29.0

Min 271.5 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 228.4 F 0.0
10:00 235.5 F 0.0
15:00 231.5 F 0.0
20:00 226.5 F 0.0
25:00 225.0 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:59
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

08/16/2023 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____