
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002559
9:01 05/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:05
STERILIZING 3:10
136.1 C 220kPa 3:10
136.2 C 223kPa 3:40
136.4 C 225kPa 4:10
136.4 C 222kPa 4:40
136.3 C 224kPa 5:10
135.9 C 221kPa 5:40
136.3 C 223kPa 6:10
136.4 C 223kPa 6:40
136.4 C 223kPa 6:40
VENTING 6:40
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
5102E5F829E16164
