
Midmark M11 - v1.0.5
Total Cycles: 2048

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 182.9 F 3.8
4:00 201.1 F 8.5
6:00 222.4 F 10.8
8:00 227.1 F 9.1
10:00 235.7 F 9.6
12:00 243.1 F 13.0
14:00 243.2 F 12.2
16:00 256.1 F 19.0
18:00 267.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.4
2:00 271.3 F 28.8
3:00 272.4 F 29.6
4:00 271.9 F 29.0

Min 271.2 F 28.7
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 236.8 F 0.0
10:00 223.2 F 0.0
15:00 215.1 F 0.0
20:00 217.3 F 0.0
25:00 218.7 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

02/14/2023 02:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____