

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1088

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2022 10:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.7 F 0.0  
2:00 161.4 F 2.9  
4:00 169.9 F 6.8  
6:00 192.3 F 11.8  
8:00 214.5 F 6.1  
10:00 229.7 F 11.5  
12:00 231.7 F 8.6  
14:00 239.4 F 10.6  
16:00 245.4 F 13.6  
18:00 257.8 F 20.3  
20:00 268.8 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.7 F 29.8  
2:00 271.4 F 29.1  
3:00 272.6 F 30.0  
4:00 272.0 F 29.3  
  
Min 271.4 F 29.1  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 206.3 F 0.0  
10:00 221.2 F 0.0  
15:00 225.0 F 0.0  
20:00 223.3 F 0.0  
25:00 221.2 F 0.0  
30:00 220.4 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 21:09  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

05/10/2022 11:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_