
Midmark M11 - v1.0.0
Total Cycles: 700

Name: MIDMARK
Sterilizer ID: V2178198
02/24/2021 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 186.4 F 3.8
4:00 209.0 F 10.3
6:00 224.3 F 6.8
8:00 241.0 F 13.5
10:00 241.7 F 11.9
12:00 240.9 F 10.9
14:00 256.2 F 18.6
16:00 269.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 271.9 F 28.3
2:00 272.4 F 28.7
3:00 272.4 F 29.0
4:00 271.5 F 28.1

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 235.2 F 0.0
10:00 227.8 F 0.0
15:00 200.3 F 0.0
20:00 191.1 F 0.0
25:00 182.6 F 0.0
30:00 194.0 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 16:57
STERILIZING: 4:00
VENTING: 2:53
DRYING: 30:00
TOTAL CYCLE: 00:56:06
CYCLE COMPLETE

02/24/2021 03:47 PM
ACCEPTED: X
REJECTED:
Operator: U01