

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4993

Name: RITTER 1  
Sterilizer ID: V2178198  
10/11/2024 03:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.9 F 0.0  
2:00 134.9 F 0.3  
4:00 157.0 F 0.7  
6:00 181.8 F 3.4  
8:00 207.0 F 7.3  
10:00 213.9 F 11.4  
12:00 226.4 F 5.9  
14:00 231.8 F 10.8  
16:00 232.1 F 7.4  
18:00 244.4 F 12.8  
20:00 241.1 F 10.9  
22:00 253.6 F 17.1  
24:00 264.4 F 23.4  
26:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.1 F 29.0  
2:00 271.8 F 28.5  
3:00 272.7 F 29.1  
4:00 271.5 F 28.7

Min 270.8 F 28.0  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 197.7 F 0.0  
10:00 192.2 F 0.0  
15:00 188.4 F 0.0  
20:00 183.8 F 0.0  
25:00 183.4 F 0.0  
30:00 189.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:00  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:04:28  
CYCLE COMPLETE

10/11/2024 04:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_