
Midmark M11 - v1.0.5
Total Cycles: 2518

Name: RITTER2
Sterilizer ID: V2313228
06/15/2023 07:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.2 F 0.0
2:00 73.5 F 0.0
4:00 78.3 F 0.0
6:00 103.2 F 0.0
8:00 148.7 F 0.5
10:00 170.5 F 3.6
12:00 195.4 F 7.1
14:00 211.4 F 10.7
16:00 226.3 F 7.3
18:00 230.1 F 9.1
20:00 238.1 F 12.6
22:00 239.6 F 10.9
24:00 235.5 F 9.0
26:00 249.6 F 15.4
28:00 260.4 F 21.6
30:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.0 F 29.2
2:00 272.6 F 29.5
3:00 272.5 F 29.8
4:00 271.6 F 29.5

Min 271.5 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 229.9 F 0.0
10:00 205.1 F 0.0
15:00 196.6 F 0.0
20:00 195.2 F 0.0
25:00 195.3 F 0.0
30:00 203.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

06/15/2023 08:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____