
Midmark M11 - v1.0.5
Total Cycles: 2243

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 07:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.9 F 0.0
2:00 89.7 F 0.0
4:00 94.7 F 0.0
6:00 109.9 F 0.1
8:00 168.2 F 0.3
10:00 196.3 F 3.4
12:00 213.7 F 7.0
14:00 224.9 F 10.8
16:00 226.4 F 7.0
18:00 234.9 F 10.0
20:00 236.7 F 9.4
22:00 243.3 F 12.6
24:00 241.5 F 11.3
26:00 253.8 F 17.6
28:00 265.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.2
2:00 271.5 F 29.0
3:00 272.3 F 29.1
4:00 271.5 F 29.0

Min 271.2 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 231.5 F 0.0
10:00 225.2 F 0.0
15:00 230.2 F 0.0
20:00 228.6 F 0.0
25:00 230.4 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 29:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:05
CYCLE COMPLETE

04/13/2023 08:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____