

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001563  
10:10 11/05/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:11  
PRESSURIZING 5:54  
STERILIZING 7:14  
136.0 C 220kPa 7:14  
136.5 C 222kPa 7:44  
136.4 C 222kPa 8:14  
136.2 C 221kPa 8:44  
136.5 C 223kPa 9:14  
136.4 C 221kPa 9:44  
136.4 C 222kPa 10:14  
136.1 C 221kPa 10:44  
VENTING 10:44  
AIR DRYING 11:51  
CYCLE COMPLETE 30:57

-----  
Digital Signature #  
13E550B15A389D09  
-----