
Midmark M11 - v1.0.5
Total Cycles: 3941

Name: RITTER2
Sterilizer ID: V2313228
09/12/2024 06:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.6 F 0.0
2:00 168.3 F 4.0
4:00 192.5 F 9.3
6:00 215.2 F 5.5
8:00 231.1 F 10.3
10:00 232.9 F 7.6
12:00 241.2 F 10.8
14:00 247.9 F 14.4
16:00 261.6 F 22.0
18:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.1 F 28.6
2:00 272.7 F 29.1
3:00 272.7 F 29.4
4:00 271.8 F 28.5

Min 271.7 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 227.5 F 0.0
10:00 212.4 F 0.0
15:00 215.2 F 0.0
20:00 222.6 F 0.0
25:00 230.7 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 18:10
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:02
CYCLE COMPLETE

09/12/2024 07:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____