
Midmark M11 - v1.0.0
Total Cycles: 1014

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 81.2 F 0.0
4:00 121.1 F 0.0
6:00 105.8 F 0.1
8:00 136.7 F 0.1
10:00 167.6 F 0.2
12:00 191.1 F 1.6
14:00 209.4 F 4.6
16:00 220.7 F 7.7
18:00 230.0 F 11.2
20:00 228.1 F 6.6
22:00 238.4 F 10.9
24:00 234.7 F 7.9
26:00 244.7 F 12.1
28:00 245.7 F 12.8
30:00 257.8 F 19.3
32:00 268.8 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.7 F 28.7
2:00 271.5 F 27.8
3:00 272.5 F 28.8
4:00 272.2 F 28.3

Min 271.4 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 226.1 F 0.0
10:00 219.3 F 0.0
15:00 208.0 F 0.0
20:00 197.2 F 0.0
25:00 201.4 F 0.0
30:00 197.5 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 33:15
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:11:09
CYCLE COMPLETE

07/15/2021 07:50 AM
ACCEPTED: X
REJECTED:
Operator: U02