
Midmark M11 - v1.0.5
Total Cycles: 2917

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 163.8 F 1.6
4:00 169.4 F 4.9
6:00 193.8 F 8.7
8:00 207.9 F 13.5
10:00 222.5 F 7.7
12:00 231.7 F 12.7
14:00 235.0 F 9.6
16:00 237.7 F 9.6
18:00 246.6 F 14.0
20:00 257.8 F 20.1
22:00 267.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.3 F 29.5
2:00 271.0 F 29.2
3:00 273.1 F 30.3
4:00 272.4 F 29.7

Min 271.0 F 28.9
Max 273.5 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 209.4 F 0.0
10:00 208.1 F 0.0
15:00 200.1 F 0.0
20:00 198.6 F 0.0
25:00 197.9 F 0.0
30:00 206.5 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:24
CYCLE COMPLETE

09/25/2023 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____