
Midmark M11 - v1.0.5
Total Cycles: 4181

Name: RITTER2
Sterilizer ID: V2313228
12/04/2024 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.4 F 0.0
2:00 165.1 F 0.4
4:00 177.6 F 3.1
6:00 202.4 F 7.6
8:00 213.5 F 12.2
10:00 226.3 F 7.7
12:00 237.1 F 12.8
14:00 237.1 F 10.8
16:00 236.4 F 9.5
18:00 250.1 F 16.0
20:00 261.4 F 22.4
22:00 271.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 273.1 F 29.9
2:00 271.5 F 29.2
3:00 272.3 F 30.0
4:00 272.4 F 29.6

Min 271.1 F 29.1
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 226.5 F 0.0
10:00 199.3 F 0.0
15:00 187.4 F 0.0
20:00 183.4 F 0.0
25:00 184.2 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

12/04/2024 12:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____