
Midmark M11 - v1.0.0
Total Cycles: 3099

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2023 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 199.7 F 3.5
4:00 214.8 F 8.6
6:00 223.9 F 9.5
8:00 231.8 F 9.7
10:00 235.1 F 8.6
12:00 244.6 F 13.3
14:00 243.7 F 12.2
16:00 256.6 F 18.9
18:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.7 F 28.3
3:00 272.4 F 29.2
4:00 272.4 F 28.9

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 221.1 F 0.0
10:00 238.9 F 0.0
15:00 227.3 F 0.0
20:00 229.0 F 0.0
25:00 228.6 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

04/12/2023 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____