

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2354

Name: RITTER 1  
Sterilizer ID: V2178198  
09/21/2022 05:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
2:00 193.8 F 4.4  
4:00 195.8 F 9.2  
6:00 218.0 F 7.6  
8:00 228.2 F 9.5  
10:00 235.6 F 9.0  
12:00 243.4 F 12.5  
14:00 241.1 F 11.0  
16:00 253.8 F 17.6  
18:00 265.1 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.4  
2:00 272.2 F 28.8  
3:00 271.6 F 28.8  
4:00 272.5 F 29.4  
  
Min 271.4 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 227.0 F 0.0  
10:00 235.8 F 0.0  
15:00 236.4 F 0.0  
20:00 231.4 F 0.0  
25:00 226.3 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:54  
CYCLE COMPLETE

09/21/2022 06:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_