

-----  
Midmark M11 - v1.0.5  
Total Cycles: 957

Name: RITTER2  
Sterilizer ID: V2313228  
04/05/2022 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.9 F 0.0  
2:00 143.8 F 0.0  
4:00 173.7 F 0.4  
6:00 202.8 F 2.9  
8:00 218.3 F 7.8  
10:00 231.7 F 12.8  
12:00 234.4 F 9.5  
14:00 238.8 F 9.5  
16:00 243.1 F 11.8  
18:00 247.1 F 14.0  
20:00 259.8 F 21.0  
22:00 271.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.7 F 29.5  
2:00 271.7 F 28.7  
3:00 271.8 F 29.3  
4:00 272.6 F 29.3

Min 271.3 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 198.5 F 0.0  
10:00 207.8 F 0.0  
15:00 219.3 F 0.0  
20:00 221.9 F 0.0  
25:00 226.4 F 0.0  
30:00 235.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 22:44  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:51  
CYCLE COMPLETE

04/05/2022 12:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_