
Midmark M11 - v1.0.5
Total Cycles: 798

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 79.4 F 0.0
4:00 112.5 F 0.1
6:00 126.5 F 0.6
8:00 156.7 F 4.1
10:00 178.6 F 7.8
12:00 205.8 F 11.7
14:00 212.2 F 5.3
16:00 225.6 F 9.5
18:00 233.0 F 10.4
20:00 238.1 F 11.3
22:00 236.8 F 9.4
24:00 250.1 F 15.9
26:00 261.4 F 22.4
28:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.4 F 30.0
2:00 272.2 F 29.4
3:00 271.3 F 29.1
4:00 272.5 F 30.0

Min 271.2 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 215.3 F 0.0
10:00 225.4 F 0.0
15:00 223.3 F 0.0
20:00 218.6 F 0.0
25:00 207.8 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 28:38
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:07:04
CYCLE COMPLETE

02/15/2022 09:07 AM
ACCEPTED: X
REJECTED:
Operator: U02